



SYMBOL	REVISION REQUIRED	DATE	AUTH.
A	Redrawn	7/14/2003	D.B.O.

Notes:

- 1. Normal tolerances +/- 0.5 mm
- 2. SENSITIVITY: 30 minutes/ 2mm
Tolerance on sensitivity is +/- 15 %.

MATERIAL: ACRYLIC
FLUID: YELLOW-GREEN



GEIER & BLUHM, INC. 594 River Street Troy, NY 12180 USA	DRAWN BY: DAVID OSTER	DATE: July 14, 2003
	Scale: NONE	DRAWING NUMBER
	CIRCULAR LEVEL VIAL	1-07537