


SYMBOL	REVISION REQUIRED	DATE	AUTH.
A	Redrawn, Revised notes	5/17/2005	D.B.O.

Notes:

- All Dimensions in inches.
- Sensitivity 10 minutes per 0.1 inch movement of the bubble.
Tolerance on sensitivity is +/- 10 %.
- Surface X** cannot be used as the level reference surface. This surface may **NOT** be perfectly flat.
- Clear Seal is not flush with respect to the .965" Ø
- MATERIAL: Glass
- FLUID: Clear Mineral Spirits



Tolerances: (except as noted)		Decimal ± .010"	Fractional ± 1/64"	Angular ± 1°
GEIER & BLUHM, INC. 594 River Street Troy, NY 12180 USA	DRAWN BY: DAVID OSTER		DATE: March 15, 2000	
	SCALE: NONE		DRAWING NUMBER	
	VIAL, CIRCULAR LEVEL		1-16100	