

SYMBOL	REVISION REQUIRED	DATE	AUTH.
A	Add Tolerances & Notes	10/63	H.F.O.
B	Redrawn, revised notes, part number was 2-6000	11/13/2003	D.B.O.
C	Flange thickness was .078"	12/31/2005	D.B.O.
D	Added Note # 4 (temperature range)	3/01/2010	D.B.O.


Notes:

- All dimensions in inches.
- SENSITIVITY: 45 minutes / .1" movement of the bubble
Tolerance on sensitivity is $\pm 10\%$.
- Part or all of the finish maybe removed from Surface "X" during calibration.
- Temperature Range: - 40° F and + 150° F

MATERIAL: Housing: Brass
Vial: Glass

FLUID: Clear Mineral Spirits

FINISH on Housing: Chrome Plate



Tolerances: (except as noted) GEIER & BLUHM, INC. 594 River Street Troy, NY 12180 USA	Decimal $\pm .010''$	Fractional $\pm 1/64''$	Angular $\pm 1^\circ$
	DRAWN BY: DAVID OSTER		DATE: November 13, 2003
	Scale: NONE		DRAWING NUMBER
MOUNTED CIRCULAR LEVEL		2-06000	