

SYMBOL	REVISION REQUIRED	DATE	AUTH.
A	Tolerances & Note 1	10/1963	H.F.O.
B	Centering Circle was .320"	06/1982	R.L.O.
C	Added Notes 2 & 3, Redrawn	09/1999	D.B.O.
D	Redrawn, revised notes, mounting holes were # 31 drill	5/28/2004	D.B.O.

Notes:

- All Dimensions in inches.
- SENSITIVITY: 10 minutes/.1" movement of bubble
Tolerance on sensitivity is +/- 10%
- Part or all of the finish maybe removed from Surface "X" during calibration
- Temperature Requirements: - 40° F to + 160° F

MATERIAL: Housing: Brass
Vial: Glass

FINISH on Housing: Chrome

FLUID: Mineral Spirits


COLOR of FLUID: Clear

TOLERANCES:

Decimal: +/- .010"

Fractional: ± 1/64"

Angular: ± ½ °



GEIER & BLUHM, INC.
594 River Street
Troy, NY 12180
USA

DRAWN BY: DAVID OSTER

Scale: NONE

**MOUNTED
CIRCULAR LEVEL**

DATE: June 10, 1963

DRAWING NUMBER

2-16000