

SYMBOL	REVISION REQUIRED	DATE	AUTH.
A	Redrawn, revised notes, Mounting Holes were # 31 Drill	12/26/2007	D.B.O.

Notes:

1. All Dimensions in inches.
2. SENSITIVITY: 10 minutes/.1" movement of bubble
Tolerance on sensitivity is +/- 10%
3. Part or all of the finish maybe removed from Surface "X" during calibration
4. Temperature Requirements: - 40° F to + 160° F

MATERIAL: Housing: Brass
Vial: Glass

FINISH on Housing: Chrome

FLUID: Mineral Spirits


COLOR of FLUID: Clear

TOLERANCES:

Decimal: +/- .010"

Fractional: ± 1/64"

Angular: ± ½ °



GEIER & BLUHM, INC. 594 River Street Troy, NY 12180 USA	DRAWN BY: DAVID OSTER	DATE: June 29, 1982
	Scale: NONE	DRAWING NUMBER 2-16017
	MOUNTED CIRCULAR LEVEL	