

SYMBOL	REVISION REQUIRED	DATE	AUTH.
A	Redrawn, revised notes, Mounting Holes were # 31 Drill	12/26/2007	D.B.O.

Notes:

1. All Dimensions in inches.
2. SENSITIVITY: 30 minutes/Division, Range $\pm 1^\circ$
Tolerance on sensitivity is $\pm 10\%$
3. Part or all of the finish maybe removed from Surface "X" during calibration

MATERIAL: Housing: Brass
Vial: Glass

FINISH on Housing: Chrome

FLUID: Mineral Spirits


COLOR of FLUID: Clear

TOLERANCES:

Decimal: $\pm .010$ "

Fractional: $\pm 1/64$ "

Angular: $\pm \frac{1}{2}^\circ$



GEIER & BLUHM, INC. 594 River Street Troy, NY 12180 USA	DRAWN BY: DAVID OSTER	DATE: June 30, 1982
	Scale: NONE	DRAWING NUMBER 2-16019
	MOUNTED CIRCULAR LEVEL	