


SYMBOL	REVISION REQUIRED	DATE	AUTH.

- Notes:**
1. All Dimensions in inches
 2. SENSITIVITY: 10 minutes per .1" movement of the bubble.
Tolerance on sensitivity is +/- 10%.
 3. Part or all of the finish maybe removed from Surface "X" during calibration
 4. MATERIAL: Housing: Brass
Vial: Glass
 5. FINISH on Housing: Black Lacquer Paint
 6. FLUID: Clear, Mineral Spirits

TOLERANCES:
 DECIMAL: $\pm .010$ "
 FRACTIONAL: $\pm 1/64$ "
 ANGULAR $\pm 1^\circ$



GEIER & BLUHM, INC. 594 River Street Troy, NY 12180 USA	DRAWN BY: DAVID OSTER	DATE: January 3, 2008
	Scale: NONE	DRAWING NUMBER 2-18050
	MOUNTED CIRCULAR LEVEL	